# Quantum Cascade Lasers for Chemical Sensing

**Frank K. Tittel*, Yury A. Bakhirkin, Anatoliy A. Kosterev, Rafal Lewicki, Stephen So, Gerard Wysocki and Robert. F. Curl**

*Rice Quantum Institute, MS 366, Rice University, 6100 Main St., Houston, TX 77005, USA*
* corresponding author: fkt@rice.edu

This talk will focus on recent advances in the development of sensors based on infrared semiconductor lasers for the detection, quantification and monitoring of trace gas species and their application to sustainable energy technologies, specifically environmental monitoring. Ultrasensitive chemical analysis of gases based on molecular absorption laser spectroscopy is a well-established technology [1,2]. The development of compact trace gas sensors based on the use of both diode lasers as well as quantum cascade (QC) and interband cascade (IC) lasers [3,4] will be emphasized. QC and IC lasers permit the targeting strong fundamental rotational-vibrational transitions in the mid-infrared. Vibrational fundamentals are one to two orders of magnitude more intense than the overtone transitions of the near infrared.

The architecture and performance of several sensitive, selective and real-time gas sensors based on near and mid-infrared semiconductor lasers will be described. To date we have detected 16 gases ($CH_4$, $H_2S$, $N_2O$, $CO_2$, CO, NO, $H_2O$, $SO_2$, $NH_3$, $C_2H_2$, OCS, $C_2H_4$, $H_2CO$, $C_2H_5OH$, $C_2HF_5$ and $CH_3COCH_3$) at the ppm to ppt level. In several cases, isotopic signatures of carbon and oxygen have also been observed. High sensitivity requires sensitivity enhancement schemes such as multipass gas absorption cells, cavity absorption enhancement techniques, or photo-acoustic absorption spectroscopy [5-7]. These methods can achieve minimum detectable absorbances in the range from $10^{-4}$ to $10^{-5}$ for field deployable gas sensors. Several recent examples of real world applications in atmospheric chemistry and the monitoring of air quality will be reported.

**References:**

[1] Rice University Laser Science Group website: http://ece.rice.edu/lasersci/
[2] R.F. Curl, F. Tittel, Annual Reports of Progress in Chemistry, Section C, **98**,217-270 (2002)
[3] F. Capasso, C. Gmachl, R. Paiella, A. Tredicucci, A. L. Hutchinson, D. L. Sivco, J. N. Baillargeon, and A. Y. Cho, "New Frontiers in Quantum Cascade Lasers and Applications", IEEE Select. Topics Quantum Electron **6**, 931-947 (2000).
[4] R.Q. Yang, J.L. Bradshaw, J.D Bruno, J.T. Pham, and D.E. Wortman, "Mid-Infrared Type-II Interband Cascade Lasers", IEEE Select. Topics Quantum Electron **38**, 559- 68 (2002).
[5] A. A. Kosterev and F. K. Tittel, "Chemical Sensors Based on Quantum Cascade Lasers", IEEE J. Quantum Electron., **38**, 582-591 (2002).
[6] F. K. Tittel, D. Richter, and A. Fried, "Mid-Infrared Laser Applications in Spectroscopy", *Solid State Mid-infrared Laser Sources* in Topics of Applied Physics, **89**, 487-559 (2003), ed. I. T. Sorokina, and K. L. Vodopoyanov, Springer-Verlag, Berlin-Heidelberg.
[7] A. A. Kosterev, F.K. Tittel, D. Serebryakov, A. Malinovsky and A. Morozov, "Applications of Quartz Tuning Fork in Spectroscopic Gas Sensing", Journal of Scientific Instruments Review **76**, 043105 (2005)